

Thyroid ELISA Kits

Product Name	Standards/ Internal control	Sample	Sample Volume	Wash	Incubation Time	Shelf Life	Sensitivity	Cat No.
T3 ELISA KIT	0-0.5-1-2.5-5-10 ng/ml CS	Serum/ plasma	50 µl	1	45+15 min RT	1 year	0.2 ng/ml	AVA- ET3-101
T4 ELISA KIT	0-2.5-5-10-15-30 ng/ml CS	Serum/ plasma	25 µl	1	45+15 min RT	1 year	1 µg/dl	AVA- ET4-102
TSH ELISA KIT	0-0.5-2.5-5-10-20-40 ng/ml CS	Serum/ plasma	50 µl	1	45+15 min RT	1 year	0.05 µIU/ml	AVA- ETS-103

Vitamin D ELISA Kits

Vit D3 elisa KIT	0-10-20-40-70-100 ng/ml	Serum/ plasma	25 micro litr	2	40+30+15 min RT	1 year	3.5 ng/ml	AVA- ED3-203
------------------	-------------------------	------------------	------------------	---	-----------------------	--------	-----------	-----------------